

## Corrigendum

No.: IIT Mandi /Admin/PUR-121 /2014-15/7368-69

Dated: 05.02.2015

1. Central Public Procurement Portal
2. IIT Mandi, Institute Website

**Sub: Corrigendum for limited Tender for: Supply Installation, Commissioning and Testing of Planetary Ball Mill**

**Ref: NIT No.: IITMANDI/Admin/PUR-121 /2014-15/7175-76 dated 24<sup>th</sup> January, 2015**

### Technical Specifications

#### 1) Sr. No.1,4,6,9 & 14 :

<b>A.</b>	<b>(Technical Specifications)</b>	
	<b>Specification Mentioned as</b>	<b>To be read as</b>
<b>1.</b>	Laboratory Planetary Ball Mill for grinding/alloying and mixing should comply to the following specifications at a maximum speed of above 1000 rpm	Laboratory Planetary Ball Mill for grinding/alloying and mixing should comply to the following specifications at speed of 1000 rpm or more
<b>4</b>	It shall be possible to use one jar of 50ml / 125 ml / 250 ml / 500 ml capacity	It shall be possible to use one jar of 50ml/80ml /250 ml / 500 ml capacity
<b>6</b>	Optionally, Vendors should quote for the jars of 50 ml, 125 ml, 250 ml and 500ml volume of Zirconium Oxide, Sintered Aluminium Oxide and stainless steel. The vendor should quote for Aeration covers and safety closing device.	Optionally, Vendors should quote for the jars of 50 ml, 125 ml, 250 ml and 500ml volume of Zirconium Oxide, Sintered Aluminium Oxide and stainless steel. The vendor should quote for safety cover and safety closing device.
<b>9</b>	The vendor should quote for Pressure and Temperature Measuring System PM Grind Control with grinding jar "comfort" 250 ml, stainless steel to record the most important thermodynamic parameters: pressure and temperature.	The vendor should quote for Pressure and Temperature Measuring System
<b>14</b>	The instrument should have Graphic display with button operations for Speed, Grinding time, Energy input, Display of drive load factor, Operating hours and intervals.	The instrument should have Graphic/digital display with button operations for Speed, Grinding time, Operating hours and other important operating parameters

All remaining technical specifications of the tender document and other Terms & condition will remain same.

**Sd/-**  
**Assistant Registrar**  
**Stores & Purchase**