



CORRIGENDUM

No. IITMANDI/S&P/PUR-17/2015-16/1038-39

Date: 13.05.2015

To

1. Institute Website IIT Mandi
2. Central Public Procurement Portal.

Sub: - Corrigendum to NIT No. IITMANDI/S&P/PUR-17/2015-16/904-05 dated 05.05.2015 regarding Supply & Installation of Integrated of Upright Ultra Deep Freezer (-80 °C) 800 Litter And Upright Deep Freezer -40c

The following changes are made to the specification of the **Upright Ultra Deep Freezer (-80 °C) 800 Litter And Upright Deep Freezer -40c**.

Annexure 1

Sr. No.	Mentioned in Tender Document	May be "Read as"
1	• Upright Ultra freezer -86°C having operating temperature of -55 to -86°C with 1°C increment. Freezer must have capacity of 570-590 litters, microprocessor controlled and password protected.	Upright Ultra freezer -86°C having operating temperature of -55 to -86°C with 1°C increment. Freezer must have capacity of 800 litters or better, microprocessor controlled and password protected.

Those who have already sent their bids as per earlier technical specification can send their revised bids in the light of the correct technical specification as mentioned above. However, they need not to submit bid fee & EMD again.

All other terms and conditions will remain the same.

Sd/-
**(Stores & Purchase)
Assistant Registrar**